

Ka-Band Waveguide Directional Coupler, 10 dB

Description:

Model SWD-1040H-28-SB is a Ka band, three-port waveguide directional coupler that delivers a 10 dB nominal coupling level and 35 dB minimum directivity across the full waveguide band from 26.5 to 40 GHz. The three-port coupler uses a traditional multi-hole and split block design to achieve a flat coupling level, high directivity, and low insertion loss. The interfaces of the



coupler are WR-28 waveguides with UG-599/U flanges. Custom coupling levels are available under different model numbers.

Features:

- Full Band Operation
- Low Insertion Loss
- High Directivity

Applications:


- Test Labs
- Instrumentation
- Sub-assemblies

Electrical Specifications:

Parameter	Minimum	Typical	Maximum
Frequency	26.5 GHz		40 GHz
Insertion Loss*		0.5 dB	
Coupling*		10 dB	
Directivity*	35 dB		
Return Loss			26 dB
Specification Temperature		+25 °C	
Operating Temperature	-40 °C		+85 °C

^{*}The definition of the insertion loss, coupling and directivity is show as following. The required termination on the waveguide port is 30 dB or better for accurate measurement.

Insertion Loss = -10
$$\log_{10} [(P2+P3)/P1]$$
 P1 Coupling Value = -10 $\log_{10} [P3/P1]$ P3





www.sagemillimeter.com | 3043 Kashiwa Street, Torrance, CA 90505 Phone: 424-757-0168 | Fax: 424-757-0188 | Email: sales@sagemillimeter.com




Ka-Band Waveguide Directional Coupler, 10 dB

Mechanical Specifications:

Item	Specification	
Through Ports	WR-28 Waveguide with UG/599 Flange	
Coupled Port	WR-28 Waveguide with UG/599 Flange	
Material	Brass	
Finish	Gold Plated	
Weight	20 Oz	
Size	6.50" (L) X 1.20" (W) x 0.90" (H)	
Outline	WD-SB-A	

Typical Coupling and Insertion Loss vs. Frequency






www.sagemillimeter.com | 3043 Kashiwa Street, Torrance, CA 90505 Phone: 424-757-0168 | Fax: 424-757-0188 | Email: sales@sagemillimeter.com



Ka-Band Waveguide Directional Coupler, 10 dB

Mechanical Outline: (Unless otherwise specified, all dimensions are in inches [millimeters])



Note:

- All data presented is collected from a sample lot. Actual data may vary unit to unit.
- The insertion loss shown includes the loss due to coupling.
- SAGE Millimeter, Inc. reserves the right to change the information presented without notice.

Caution:

Any foreign objects in the waveguide will degrade performance and/or damage the device.





www.sagemillimeter.com | 3043 Kashiwa Street, Torrance, CA 90505 Phone: 424-757-0168 | Fax: 424-757-0188 | Email: sales@sagemillimeter.com